



- Integrated safety mechanism
- Minimizes exposure to the contaminated needle
- Allows for one-handed activation
- Once activated, the needle is safely retracted through disposal
- Easy to use

**THE
NEEDLE THAT
DISAPPEARS**

VanishPoint® IV Catheters 50 per box/200 per case

Catalog Number	Gauge	Length	Catheter Material	Color Code
31221	24	¾"	Radiopaque PUR	Yellow
35221	24	¾"	Radiopaque FEP*	Yellow
31331	22	1"	Radiopaque PUR	Blue
35331	22	1"	Radiopaque FEP*	Blue
31441	20	1¼"	Radiopaque PUR	Pink
35441	20	1¼"	Radiopaque FEP*	Pink
31541	18	1¼"	Radiopaque PUR	Green
35541	18	1¼"	Radiopaque FEP*	Green
31641	16	1¼"	Radiopaque PUR	Grey
35641	16	1¼"	Radiopaque FEP*	Grey
31741	14	1¼"	Radiopaque PUR	Orange
35741	14	1¼"	Radiopaque FEP*	Orange

18-24 gauge IV catheters are suitable for use with power injectors up to 325 psi

**Special order items - please contact RTI for ordering information.*

Other Available Safety Products



VanishPoint®
Syringe



EasyPoint®
Retractable
Needle



VanishPoint®
Blood Collection
Set



EasyPoint® Blood
Collection Tube
Holder with needle

To order these products,
please contact
rtisales@retractable.com



RETRACTABLE
TECHNOLOGIES, INC.®

© 2012, 2025
Retractable Technologies, Inc.
All rights reserved MKT-BRO-
VPVC-0003 Rev.0 EN, AUG.2025

511 Lobo Lane, Little Elm
Texas 75068-5295 USA
Tel.: (972) 294-1770
Fax: (972) 294-4400
Toll Free: 1-888-703-1010



**The New Standard
for Safety**



**SAVES MONEY,
TIME AND
LIVES**



www.retractable.com



RETRACTABLE
TECHNOLOGIES, INC.®



VanishPoint[®]
IV CATHETER

SAFE AND EFFECTIVE

ADVANTAGE

A safety IV Catheter featuring the patented VanishPoint[®] automated retraction!

Sharp introducer needle for insertion ease and patient comfort



Clear finger grips and flashback chamber allow for handling ease and flashback visualization

Push-off tab facilitates one handed advancement and indicates bevel orientation



Color coded on both catheter hub and housing



Automatic needle retraction occurs with slight depression of the needle housing. Once activated, the needle is safely retracted into the housing, reducing the risk of needlestick injuries



1 Perform venipuncture observing for blood return. Advance the catheter slightly to assure full catheter entry into the vein.



2 Holding the flashback chamber stationary, disengage the hub from the introducer needle and advance the catheter into the vein.



3 Before withdrawing the introducer needle from the catheter, depress the color coded end of the needle housing to activate the automated retraction.



4 The introducer needle will retract into the housing. Dispose of the introducer needle in an appropriate sharps container.